



UTENSILI PER PROFESSIONISTI DAL 1913

DICHIARAZIONE DI CONFORMITÀ CE

del 13/11/2019

Multimetri digitali a pinza

H 5440/1

13_11_2019

ABC Tools S.p.A.

Viale Europa, 68/70 - 20093 Cologno Monzese (MI) - Italia
T. +39 02 251111.1 - F. +39 02 2538379 - info@abctools.it



EU DECLARATION OF CONFORMITY

Product Identification:

Amprobe ACD-10 PLUS, ACD-10 PLUS TRMS - Clamp Meter

First declared: 2009

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO 9001:2015 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

- Directive 2011/65/EU, RoHS**
- Directive 2014/30/EU, Electromagnetic Compatibility (EMC)**
- Directive 2014/35/EU, Low Voltage (LVD)**

Standards Used:

- EN 50581: 2012 Restriction of the use of hazardous substances
 - EN 61326-1: 2013 Electrical equipment for measurement, control and laboratory-EMC; General
 - EN 61010-1: 2010 Safety requirements for measurement, control and laboratory; General
 - EN 61010-2-032: 2012 Particular requirements for hand-held and hand manipulated current sensors
 - EN 61010-2-033: 2012 Particular requirements for hand-held Multimeters for measuring Mains voltage
-

Special Conditions:

None

Manufacturer

Amprobe
6920 Seaway Boulevard M/S 143F
Everett, WA 98203; USA

DC20191113132922

Thomas M. Smith P.E.

Product Compliance Manager, Fluke Corporation

DC20191113132922 - This declaration supersedes all previous declarations for this product.

13 November 2019